
Midmark M11 - v1.0.0
Total Cycles: 2246

Name: RITTER 1
Sterilizer ID: V2178198
08/25/2022 09:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.3 F 0.0
2:00 178.8 F 0.3
4:00 192.7 F 1.3
6:00 215.1 F 4.6
8:00 228.0 F 9.4
10:00 237.0 F 9.6
12:00 241.0 F 11.5
14:00 238.3 F 9.5
16:00 238.6 F 8.9
18:00 250.4 F 15.1
20:00 262.3 F 21.9
22:00 272.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.0
1:00 272.6 F 28.8
2:00 271.5 F 27.8
3:00 272.8 F 28.7
4:00 271.5 F 27.9

Min 271.4 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 243.2 F 0.0
10:00 229.4 F 0.0
15:00 228.1 F 0.0
20:00 227.3 F 0.0
25:00 231.6 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 22:12
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:27
CYCLE COMPLETE

08/25/2022 10:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____