

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2738

Name: RITTER 1  
Sterilizer ID: V2178198  
01/04/2023 05:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.1 F 0.0  
2:00 173.3 F 0.0  
4:00 192.4 F 3.6  
6:00 211.6 F 8.6  
8:00 226.4 F 11.7  
10:00 233.0 F 9.6  
12:00 236.1 F 8.9  
14:00 245.4 F 12.9  
16:00 245.2 F 12.9  
18:00 258.8 F 20.2  
20:00 270.8 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.2 F 28.8  
2:00 272.0 F 28.9  
3:00 271.9 F 28.6  
4:00 272.3 F 29.1

Min 271.4 F 28.2  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.9 F 0.0  
5:00 220.1 F 0.0  
10:00 240.8 F 0.0  
15:00 245.1 F 0.0  
20:00 236.1 F 0.0  
25:00 220.8 F 0.0  
30:00 215.0 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 20:43  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:58:48  
CYCLE COMPLETE

01/04/2023 05:58 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_