
Midmark M11 - v1.0.5
Total Cycles: 842

Name: RITTER2
Sterilizer ID: V2313228
02/28/2022 10:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.2 F 0.0
2:00 103.4 F 0.8
4:00 122.4 F 2.7
6:00 152.8 F 5.8
8:00 180.3 F 9.9
10:00 204.9 F 6.8
12:00 216.9 F 8.3
14:00 228.9 F 12.9
16:00 232.6 F 9.1
18:00 238.7 F 10.3
20:00 244.8 F 13.2
22:00 257.0 F 19.7
24:00 267.9 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.7 F 30.0
2:00 271.7 F 29.4
3:00 271.2 F 29.4
4:00 272.6 F 30.4

Min 270.9 F 29.0
Max 272.8 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.6 F 0.0
5:00 233.8 F 0.0
10:00 194.5 F 0.0
15:00 192.1 F 0.0
20:00 190.8 F 0.0
25:00 206.4 F 0.0
30:00 202.1 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:22
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:03:25
CYCLE COMPLETE

02/28/2022 11:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____