

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2219

Name: RITTER 1  
Sterilizer ID: V2178198  
08/15/2022 07:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.9 F 0.0  
2:00 72.5 F 0.0  
4:00 77.6 F 0.0  
6:00 97.8 F 0.2  
8:00 134.3 F 0.5  
10:00 168.7 F 3.2  
12:00 185.7 F 6.4  
14:00 200.3 F 9.7  
16:00 212.8 F 13.2  
18:00 220.9 F 6.2  
20:00 231.2 F 10.0  
22:00 238.0 F 10.0  
24:00 240.1 F 11.0  
26:00 235.8 F 8.5  
28:00 249.2 F 14.7  
30:00 259.7 F 20.6  
32:00 269.4 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.5 F 28.9  
2:00 271.5 F 28.1  
3:00 272.1 F 28.8  
4:00 272.1 F 28.6  
  
Min 271.2 F 28.0  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.1 F 0.0  
5:00 233.7 F 0.0  
10:00 218.1 F 0.0  
15:00 212.3 F 0.0  
20:00 231.4 F 0.0  
25:00 229.5 F 0.0  
30:00 229.2 F 0.0  
  
Drying Complete

FILLING: 0:45  
HEATING: 33:07  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:11:03  
CYCLE COMPLETE

08/15/2022 08:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_