
Midmark M11 - v1.0.5
Total Cycles: 1218

Name: RITTER2
Sterilizer ID: V2313228
06/24/2022 09:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.9 F 0.0
2:00 180.1 F 3.0
4:00 189.5 F 7.2
6:00 214.4 F 12.1
8:00 222.7 F 7.4
10:00 234.8 F 12.8
12:00 236.4 F 10.3
14:00 235.3 F 8.4
16:00 249.0 F 15.3
18:00 260.4 F 21.8
20:00 271.2 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.2 F 29.6
2:00 271.3 F 29.2
3:00 272.1 F 29.6
4:00 271.3 F 29.0

Min 271.2 F 29.0
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 218.7 F 0.0
5:00 224.4 F 0.0
10:00 229.4 F 0.0
15:00 225.5 F 0.0
20:00 220.9 F 0.0
25:00 218.0 F 0.0
30:00 215.1 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 20:40
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:36
CYCLE COMPLETE

06/24/2022 10:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____