
Midmark M11 - v1.0.5
Total Cycles: 466

Name: RITTER2
Sterilizer ID: V2313228
10/21/2021 08:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.6 F 0.0
2:00 137.4 F 1.5
4:00 147.7 F 4.6
6:00 178.4 F 9.2
8:00 203.9 F 8.4
10:00 221.5 F 8.3
12:00 235.8 F 12.6
14:00 238.4 F 11.2
16:00 237.7 F 9.9
18:00 251.7 F 16.8
20:00 263.7 F 23.8
22:00 271.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.7 F 29.7
2:00 271.5 F 29.1
3:00 272.6 F 29.9
4:00 271.9 F 29.1

Min 271.5 F 28.9
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 223.0 F 0.0
10:00 226.3 F 0.0
15:00 224.6 F 0.0
20:00 221.4 F 0.0
25:00 218.3 F 0.0
30:00 217.7 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 22:01
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

10/21/2021 09:10 AM
ACCEPTED: X
REJECTED:
Operator: U02