
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001715
9:54 21/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:43
PRESSURIZING 0:53
STERILIZING 2:02
136.4 C 224kPa 2:02
136.3 C 224kPa 2:32
136.2 C 221kPa 3:02
136.4 C 224kPa 3:32
136.4 C 224kPa 4:02
136.3 C 223kPa 4:32
136.4 C 224kPa 5:02
136.4 C 224kPa 5:32
136.4 C 224kPa 5:32
VENTING 5:33
AIR DRYING 6:22
CYCLE COMPLETE 26:37

Digital Signature #
970844661670A1CA
