
Midmark M11 - v1.0.0
Total Cycles: 3927

Name: RITTER 1
Sterilizer ID: V2178198
10/20/2023 07:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.3 F 0.0
2:00 72.9 F 0.0
4:00 88.0 F 0.0
6:00 117.3 F 0.1
8:00 146.6 F 0.6
10:00 176.1 F 3.9
12:00 207.1 F 7.1
14:00 226.0 F 10.4
16:00 231.8 F 8.9
18:00 226.9 F 7.6
20:00 238.5 F 11.6
22:00 233.6 F 8.1
24:00 245.4 F 12.9
26:00 243.2 F 11.9
28:00 255.0 F 18.1
30:00 265.6 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.5 F 29.1
2:00 271.2 F 28.3
3:00 272.3 F 29.2
4:00 271.6 F 28.4

Min 271.1 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.5 F 0.0
5:00 250.6 F 0.0
10:00 229.4 F 0.0
15:00 225.0 F 0.0
20:00 225.8 F 0.0
25:00 230.3 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 31:51
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:09:51
CYCLE COMPLETE

10/20/2023 09:05 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____