

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 003969  
9:35 08/07/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:50  
PRESSURIZING 3:51  
STERILIZING 4:57  
136.4 C 222kPa 4:57  
136.4 C 224kPa 5:27  
136.4 C 224kPa 5:57  
136.4 C 223kPa 6:27  
136.4 C 223kPa 6:57  
136.3 C 222kPa 7:27  
136.4 C 222kPa 7:57  
136.3 C 221kPa 8:27  
136.3 C 221kPa 8:27  
VENTING 8:27  
AIR DRYING 9:20  
CYCLE COMPLETE 10:24

-----  
Digital Signature #  
FD4296006B374494  
-----