

-----  
Midmark M11 - v1.0.0  
Total Cycles: 843

Name: RITTER 1  
Sterilizer ID: V2178198  
05/27/2021 10:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.7 F 0.0  
2:00 176.8 F 3.6  
4:00 194.3 F 8.7  
6:00 214.5 F 8.0  
8:00 220.7 F 10.1  
10:00 229.2 F 7.2  
12:00 240.9 F 13.4  
14:00 243.7 F 11.8  
16:00 256.7 F 18.7  
18:00 267.8 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 271.9 F 29.0  
2:00 272.5 F 29.0  
3:00 272.1 F 29.0  
4:00 272.1 F 28.9  
  
Min 271.1 F 28.5  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.2 F 0.0  
5:00 226.6 F 0.0  
10:00 223.5 F 0.0  
15:00 212.2 F 0.0  
Dry Cycle Aborted

FILLING: 1:36  
HEATING: 19:08  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 19:30  
TOTAL CYCLE: 00:47:15

05/27/2021 11:32 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_