

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000096  
8:43 02/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:18  
PRESSURIZING 3:24  
STERILIZING 4:59  
136.0 C 220kPa 4:59  
136.4 C 225kPa 5:29  
136.4 C 226kPa 5:59  
136.4 C 226kPa 6:29  
136.4 C 226kPa 6:59  
136.3 C 224kPa 7:29  
136.4 C 226kPa 7:59  
136.4 C 225kPa 8:29  
136.4 C 225kPa 8:29  
VENTING 8:29  
AIR DRYING 9:31  
CYCLE COMPLETE 24:31  
-----

Digital Signature #  
6DC189F882B27F38  
-----