
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001136
8:30 23/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:36
PRESSURIZING 4:37
STERILIZING 6:01
136.0 C 220kPa 6:01
136.4 C 225kPa 6:31
136.1 C 221kPa 7:01
136.2 C 222kPa 7:31
136.4 C 222kPa 8:01
136.3 C 224kPa 8:31
136.4 C 225kPa 9:01
136.4 C 224kPa 9:31
136.4 C 224kPa 9:31
VENTING 9:31
AIR DRYING 10:33
CYCLE COMPLETE 21:09

Digital Signature #
F711A773E5EEEA98
