
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001797
10:46 19/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:07
STERILIZING 3:09
136.0 C 220kPa 3:09
136.2 C 222kPa 3:39
136.4 C 224kPa 4:09
136.0 C 222kPa 4:39
136.2 C 224kPa 5:09
136.4 C 223kPa 5:39
136.4 C 224kPa 6:09
136.4 C 224kPa 6:39
136.4 C 224kPa 6:39
VENTING 6:39
AIR DRYING 7:32
CYCLE COMPLETE 7:58

Digital Signature #
1308FBA385EAF91
