
Midmark M11 - v1.0.5
Total Cycles: 1672

Name: RITTER2
Sterilizer ID: V2313228
11/02/2022 04:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.9 F 0.0
2:00 156.8 F 2.6
4:00 174.4 F 6.1
6:00 195.1 F 10.8
8:00 218.1 F 5.2
10:00 226.8 F 10.9
12:00 228.7 F 7.3
14:00 243.1 F 13.2
16:00 242.1 F 11.8
18:00 254.5 F 18.2
20:00 265.2 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.2
1:00 271.7 F 29.7
2:00 272.2 F 29.7
3:00 271.5 F 29.1
4:00 271.4 F 29.6

Min 271.2 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.1 F 0.0
5:00 212.7 F 0.0
10:00 234.2 F 0.0
15:00 247.7 F 0.0
20:00 233.3 F 0.0
25:00 222.8 F 0.0
30:00 217.2 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:23
CYCLE COMPLETE

11/02/2022 05:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____