
Midmark M11 - v1.0.0
Total Cycles: 1260

Name: RITTER 1
Sterilizer ID: V2178198
10/13/2021 09:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.7 F 0.0
2:00 196.3 F 3.2
4:00 196.0 F 8.3
6:00 236.1 F 8.7
8:00 234.1 F 10.4
10:00 233.6 F 7.3
12:00 243.2 F 11.3
14:00 246.1 F 13.0
16:00 258.8 F 19.9
18:00 270.0 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 271.5 F 28.6
2:00 271.4 F 28.6
3:00 271.5 F 28.5
4:00 271.8 F 28.6

Min 271.3 F 28.5
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.9 F 0.0
5:00 238.3 F 0.0
10:00 226.8 F 0.0
15:00 229.8 F 0.0
20:00 251.8 F 0.0
25:00 231.4 F 0.0
30:00 205.0 F 0.0

Drying Complete

FILLING: 1:56
HEATING: 19:32
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

10/13/2021 10:56 AM
ACCEPTED: X
REJECTED:
Operator: U01