
Midmark M11 - v1.0.5
Total Cycles: 2730

Name: RITTER2
Sterilizer ID: V2313228
08/03/2023 01:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.1 F 0.0
2:00 124.2 F 0.0
4:00 148.4 F 0.5
6:00 178.7 F 4.3
8:00 196.6 F 8.7
10:00 210.2 F 13.4
12:00 224.4 F 7.9
14:00 235.8 F 12.7
16:00 235.9 F 9.8
18:00 236.8 F 9.2
20:00 247.6 F 14.7
22:00 259.4 F 21.3
24:00 270.4 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.7 F 30.0
2:00 271.9 F 29.2
3:00 271.5 F 29.5
4:00 272.9 F 29.9

Min 271.1 F 29.0
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 216.0 F 0.0
10:00 217.4 F 0.0
15:00 220.4 F 0.0
20:00 222.5 F 0.0
25:00 215.6 F 0.0
30:00 219.9 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:53
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:03:04
CYCLE COMPLETE

08/03/2023 02:48 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____