

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1300

Name: RITTER2  
Sterilizer ID: V2313228  
07/14/2022 01:53 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 149.2 F 0.0  
2:00 157.7 F 0.2  
4:00 187.4 F 2.8  
6:00 207.9 F 7.6  
8:00 223.2 F 12.9  
10:00 230.2 F 8.6  
12:00 240.1 F 11.9  
14:00 242.1 F 12.2  
16:00 241.2 F 11.3  
18:00 254.4 F 18.0  
20:00 265.9 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.4 F 29.1  
2:00 272.3 F 29.4  
3:00 272.1 F 28.8  
4:00 272.6 F 29.5  
  
Min 271.6 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.3 F 0.0  
5:00 182.4 F 0.0  
10:00 167.7 F 0.0  
15:00 158.0 F 0.0  
20:00 149.0 F 0.0  
25:00 144.3 F 0.0  
30:00 143.9 F 0.0  
Drying Complete

FILLING: 1:42  
HEATING: 21:41  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:34  
CYCLE COMPLETE

07/14/2022 02:53 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_