

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1410

Name: RITTER2  
Sterilizer ID: V2313228  
08/24/2022 04:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 134.1 F 0.0  
2:00 139.2 F 0.0  
4:00 169.1 F 2.0  
6:00 193.6 F 5.1  
8:00 210.9 F 9.5  
10:00 223.9 F 8.5  
12:00 232.5 F 10.2  
14:00 233.6 F 8.5  
16:00 242.6 F 11.9  
18:00 245.4 F 13.4  
20:00 258.0 F 20.2  
22:00 269.4 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.0  
1:00 272.1 F 29.2  
2:00 272.5 F 29.7  
3:00 271.6 F 28.9  
4:00 272.4 F 29.5

Min 271.6 F 28.9  
Max 272.5 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.0 F 0.0  
5:00 207.7 F 0.0  
10:00 233.4 F 0.0  
15:00 233.9 F 0.0  
20:00 222.7 F 0.0  
25:00 226.5 F 0.0  
30:00 228.0 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 22:57  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:01:08  
CYCLE COMPLETE

08/24/2022 05:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_