
Midmark M11 - v1.0.0
Total Cycles: 1620

Name: RITTER 1
Sterilizer ID: V2178198
02/10/2022 04:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.4 F 0.0
2:00 169.0 F 2.8
4:00 187.8 F 7.0
6:00 208.6 F 11.8
8:00 222.2 F 5.9
10:00 235.5 F 10.8
12:00 231.2 F 7.1
14:00 245.2 F 13.4
16:00 243.4 F 12.1
18:00 255.8 F 18.5
20:00 267.3 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.7 F 28.9
2:00 271.8 F 28.7
3:00 272.3 F 28.9
4:00 271.6 F 28.6

Min 271.5 F 28.4
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.1 F 0.0
5:00 208.4 F 0.0
10:00 193.3 F 0.0
15:00 191.5 F 0.0
20:00 188.9 F 0.0
25:00 186.7 F 0.0
30:00 189.5 F 0.0
Drying Complete

FILLING: 2:18
HEATING: 21:25
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:01:08
CYCLE COMPLETE

02/10/2022 05:40 PM
ACCEPTED: X
REJECTED:
Operator: U02