
Midmark M11 - v1.0.5
Total Cycles: 3612

Name: RITTER2
Sterilizer ID: V2313228
05/13/2024 03:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.7 F 0.0
2:00 130.7 F 1.7
4:00 149.8 F 5.2
6:00 178.2 F 9.5
8:00 199.7 F 6.7
10:00 217.1 F 7.7
12:00 228.5 F 12.7
14:00 233.2 F 9.6
16:00 236.1 F 9.4
18:00 245.5 F 13.8
20:00 257.7 F 20.4
22:00 268.9 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.6 F 29.9
2:00 271.4 F 29.2
3:00 272.7 F 30.1
4:00 271.9 F 29.3

Min 271.3 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.9 F 0.0
5:00 212.0 F 0.0
10:00 219.6 F 0.0
15:00 233.4 F 0.0
20:00 230.0 F 0.0
25:00 225.3 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 23:08
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:01:20
CYCLE COMPLETE

05/13/2024 04:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____