

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002608  
7:25 03/11/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:22  
PRESSURIZING 1:24  
STERILIZING 2:25  
136.4 C 222kPa 2:25  
136.4 C 222kPa 2:55  
136.1 C 222kPa 3:25  
136.3 C 221kPa 3:55  
136.2 C 220kPa 4:25  
136.4 C 223kPa 4:55  
136.2 C 219kPa 5:25  
135.9 C 220kPa 5:55  
135.9 C 220kPa 5:55  
VENTING 5:55  
AIR DRYING 6:45  
CYCLE COMPLETE 6:46

-----  
Digital Signature #  
85DDEC23C2EF7DE2  
-----