
Midmark M11 - v1.0.0
Total Cycles: 3138

Name: RITTER 1
Sterilizer ID: V2178198
04/20/2023 12:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.0 F 0.0
2:00 197.0 F 0.0
4:00 205.0 F 4.4
6:00 220.9 F 9.4
8:00 234.6 F 8.8
10:00 240.7 F 12.1
12:00 239.8 F 10.6
14:00 238.7 F 9.7
16:00 253.8 F 17.1
18:00 266.1 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.1
1:00 272.4 F 28.7
2:00 271.4 F 28.0
3:00 272.5 F 28.8
4:00 271.7 F 28.2

Min 271.3 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 212.9 F 0.0
10:00 222.9 F 0.0
15:00 226.9 F 0.0
20:00 228.5 F 0.0
25:00 230.1 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 19:34
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:58:10
CYCLE COMPLETE

04/20/2023 01:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____