
Midmark M11 - v1.0.5
Total Cycles: 2408

Name: RITTER2
Sterilizer ID: V2313228
05/24/2023 12:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.7 F 0.0
2:00 154.8 F 2.9
4:00 175.5 F 6.7
6:00 197.5 F 11.7
8:00 214.7 F 6.1
10:00 230.6 F 11.5
12:00 232.4 F 8.8
14:00 238.9 F 10.1
16:00 246.9 F 14.0
18:00 259.0 F 20.7
20:00 270.0 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.7
1:00 272.9 F 29.8
2:00 272.9 F 30.2
3:00 272.1 F 30.1
4:00 272.7 F 29.8

Min 271.9 F 29.6
Max 273.4 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.5 F 0.0
5:00 231.7 F 0.0
10:00 232.6 F 0.0
15:00 229.8 F 0.0
20:00 232.0 F 0.0
25:00 221.6 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 21:37
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:60
CYCLE COMPLETE

05/24/2023 01:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____