

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000811  
9:42 01/12/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:32  
PRESSURIZING 3:36  
STERILIZING 5:10  
136.0 C 228kPa 5:10  
136.4 C 232kPa 5:40  
136.1 C 230kPa 6:10  
136.3 C 232kPa 6:40  
136.4 C 233kPa 7:10  
136.4 C 231kPa 7:40  
136.4 C 230kPa 8:10  
136.4 C 232kPa 8:40  
136.4 C 232kPa 8:40  
VENTING 8:40  
AIR DRYING 9:37  
CYCLE COMPLETE 24:37  
-----

Digital Signature #  
D45D9FA7DF1324E7  
-----