
Midmark M11 - v1.0.0
Total Cycles: 2158

Name: RITTER 1
Sterilizer ID: V2178198
07/20/2022 02:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.2 F 0.0
2:00 175.7 F 3.7
4:00 189.6 F 8.1
6:00 213.9 F 13.5
8:00 225.7 F 7.2
10:00 240.1 F 12.1
12:00 236.5 F 9.2
14:00 238.4 F 9.6
16:00 247.7 F 14.0
18:00 259.8 F 20.6
20:00 270.7 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.5 F 28.8
2:00 271.3 F 28.1
3:00 272.5 F 29.0
4:00 272.0 F 28.3

Min 271.3 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.3 F 0.0
5:00 227.7 F 0.0
10:00 220.6 F 0.0
15:00 207.7 F 0.0
20:00 207.8 F 0.0
25:00 223.1 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 20:48
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:21
CYCLE COMPLETE

07/20/2022 03:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____