

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2080

Name: RITTER 1  
Sterilizer ID: V2178198  
07/04/2022 09:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.5 F 0.0  
2:00 137.5 F 0.1  
4:00 162.1 F 1.3  
6:00 182.1 F 4.3  
8:00 197.2 F 8.1  
10:00 213.5 F 12.5  
12:00 222.0 F 6.1  
14:00 233.7 F 10.7  
16:00 232.9 F 7.4  
18:00 244.6 F 13.1  
20:00 242.9 F 11.5  
22:00 255.3 F 18.0  
24:00 266.4 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.4 F 28.7  
2:00 271.4 F 28.2  
3:00 272.4 F 28.8  
4:00 271.5 F 28.1

Min 271.4 F 28.0  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.2 F 0.0  
5:00 217.8 F 0.0  
10:00 228.2 F 0.0  
15:00 232.0 F 0.0  
20:00 229.6 F 0.0  
25:00 225.5 F 0.0  
30:00 224.8 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 25:36  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:03:34  
CYCLE COMPLETE

07/04/2022 10:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_