
Midmark M11 - v1.0.0
Total Cycles: 3813

Name: RITTER 1
Sterilizer ID: V2178198
09/25/2023 04:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.0 F 0.0
2:00 178.5 F 2.0
4:00 187.7 F 5.8
6:00 203.4 F 10.3
8:00 221.3 F 7.7
10:00 228.3 F 9.2
12:00 237.5 F 10.9
14:00 239.4 F 11.2
16:00 238.1 F 9.6
18:00 251.3 F 15.9
20:00 262.4 F 22.3
22:00 272.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.2 F 28.7
2:00 272.0 F 29.0
3:00 271.8 F 28.4
4:00 272.3 F 29.1

Min 271.3 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 219.9 F 0.0
10:00 228.7 F 0.0
15:00 231.5 F 0.0
20:00 226.6 F 0.0
25:00 226.7 F 0.0
30:00 230.8 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 22:22
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:01:46
CYCLE COMPLETE

09/25/2023 05:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____