
Midmark M11 - v1.0.5
Total Cycles: 3449

Name: RITTER2
Sterilizer ID: V2313228
03/11/2024 09:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.7 F 0.0
2:00 69.6 F 0.0
4:00 72.7 F 0.0
6:00 97.0 F 0.8
8:00 138.3 F 3.3
10:00 166.3 F 6.3
12:00 185.3 F 9.8
14:00 202.8 F 11.1
16:00 211.7 F 6.1
18:00 223.0 F 9.9
20:00 234.2 F 10.5
22:00 238.0 F 10.5
24:00 235.6 F 8.7
26:00 246.8 F 14.1
28:00 257.7 F 20.1
30:00 267.8 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.0 F 29.3
2:00 272.2 F 29.8
3:00 271.2 F 29.0
4:00 272.1 F 29.5

Min 271.2 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.4 F 0.0
5:00 231.7 F 0.0
10:00 217.4 F 0.0
15:00 181.0 F 0.0
20:00 176.6 F 0.0
25:00 174.6 F 0.0
30:00 173.4 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:23
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:09:33
CYCLE COMPLETE

03/11/2024 10:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____