
Midmark M11 - v1.0.0
Total Cycles: 4288

Name: RITTER 1
Sterilizer ID: V2178198
02/09/2024 10:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.3 F 0.0
2:00 141.9 F 1.3
4:00 141.0 F 3.7
6:00 162.6 F 6.6
8:00 188.8 F 10.4
10:00 211.5 F 5.4
12:00 214.5 F 6.4
14:00 224.0 F 10.3
16:00 230.0 F 8.3
18:00 238.0 F 10.9
20:00 234.5 F 8.4
22:00 248.0 F 14.5
24:00 258.7 F 20.3
26:00 268.4 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.3 F 29.3
2:00 272.1 F 28.9
3:00 271.8 F 28.8
4:00 272.8 F 29.7

Min 271.0 F 28.4
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 213.5 F 0.0
10:00 239.9 F 0.0
15:00 243.8 F 0.0
20:00 229.3 F 0.0
25:00 228.9 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 27:22
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:06:08
CYCLE COMPLETE

02/09/2024 11:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____