
Midmark M11 - v1.0.0
Total Cycles: 675

Name: MIDMARK
Sterilizer ID: V2178198
02/18/2021 08:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.0 F 0.0
2:00 73.9 F 0.0
4:00 82.2 F 0.1
6:00 113.9 F 0.1
8:00 152.6 F 0.2
10:00 184.0 F 0.3
12:00 202.9 F 2.8
14:00 218.6 F 5.7
16:00 226.8 F 8.7
18:00 235.0 F 12.4
20:00 235.2 F 8.5
22:00 245.0 F 11.8
24:00 244.0 F 11.9
26:00 241.2 F 10.4
28:00 254.5 F 17.0
30:00 265.7 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 27.6
1:00 272.5 F 28.3
2:00 271.1 F 27.6
3:00 272.7 F 28.3
4:00 271.4 F 27.5

Min 271.1 F 27.5
Max 272.7 F 28.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.0 F 0.0
5:00 233.9 F 0.0
10:00 228.2 F 0.0
15:00 216.9 F 0.0
20:00 213.7 F 0.0
25:00 199.7 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 31:44
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:09:51
CYCLE COMPLETE

02/18/2021 09:20 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____