
Midmark M11 - v1.0.0
Total Cycles: 1773

Name: RITTER 1
Sterilizer ID: V2178198
03/31/2022 06:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.6 F 0.0
2:00 80.0 F 0.0
4:00 105.4 F 0.0
6:00 134.9 F 0.2
8:00 160.1 F 1.7
10:00 180.7 F 4.2

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 188F
CHAMBER PRESS 5.7PSI
** ITEMS NOT STERILE **

03/31/2022 07:05 AM