

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1652

Name: RITTER 1  
Sterilizer ID: V2178198  
02/22/2022 10:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.7 F 0.0  
2:00 88.2 F 0.0  
4:00 121.2 F 0.5  
6:00 154.4 F 2.2  
8:00 182.5 F 4.9  
10:00 201.6 F 8.5  
12:00 218.9 F 12.7  
14:00 221.3 F 5.5  
16:00 233.3 F 9.9  
18:00 234.1 F 8.3  
20:00 242.7 F 11.8  
22:00 239.1 F 9.7  
24:00 254.5 F 17.4  
26:00 267.3 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.0  
1:00 272.6 F 28.7  
2:00 272.3 F 28.2  
3:00 271.6 F 28.0  
4:00 272.3 F 28.6

Min 271.6 F 27.8  
Max 272.8 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
Dry Cycle Aborted

FILLING: 1:14  
HEATING: 27:08  
STERILIZING: 4:00  
VENTING: 2:28  
DRYING: 0:06  
TOTAL CYCLE: 00:34:56

02/22/2022 10:39 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_