
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001696
7:31 08/06/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:41
STERILIZING 4:46
136.4 C 221kPa 4:46
136.3 C 219kPa 5:16
136.2 C 219kPa 5:46
136.4 C 224kPa 6:16
136.2 C 223kPa 6:46
136.4 C 222kPa 7:16
136.4 C 222kPa 7:46
136.5 C 224kPa 8:16
136.5 C 224kPa 8:16
VENTING 8:16
AIR DRYING 9:11
CYCLE COMPLETE 9:22

Digital Signature #
D433FBBD7504BB2B
