

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3420

Name: RITTER 1  
Sterilizer ID: V2178198  
06/21/2023 09:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 205.5 F 0.0  
2:00 199.1 F 2.8  
4:00 200.2 F 7.2  
6:00 218.7 F 12.8  
8:00 228.6 F 8.1  
10:00 239.9 F 12.6  
12:00 241.0 F 11.3  
14:00 238.9 F 9.7  
16:00 252.8 F 16.5  
18:00 264.4 F 23.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.6 F 28.9  
2:00 271.8 F 28.2  
3:00 272.0 F 28.7  
4:00 272.5 F 28.7  
  
Min 271.3 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.1 F 0.0  
5:00 232.0 F 0.0  
10:00 229.7 F 0.0  
15:00 230.3 F 0.0  
20:00 226.3 F 0.0  
25:00 228.5 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 19:58  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:48  
CYCLE COMPLETE

06/21/2023 10:30 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_