
Midmark M11 - v1.0.5
Total Cycles: 240

Name: RITTER2
Sterilizer ID: V2313228
07/22/2021 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.1 F 0.0
2:00 175.4 F 3.5
4:00 194.7 F 8.8
6:00 217.6 F 8.7
8:00 231.8 F 10.3
10:00 231.0 F 7.8
12:00 241.2 F 11.6
14:00 245.9 F 13.5
16:00 258.4 F 20.5
18:00 269.6 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.3
1:00 272.5 F 29.5
2:00 272.9 F 30.0
3:00 271.8 F 29.3
4:00 272.7 F 29.7

Min 271.8 F 29.1
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.9 F 0.0
Dry Cycle Aborted

FILLING: 2:47
HEATING: 18:56
STERILIZING: 4:00
VENTING: 3:07
DRYING: 4:47
TOTAL CYCLE: 00:33:37

07/22/2021 01:45 PM
ACCEPTED: X
REJECTED:
Operator: U02