

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005830  
13:42 23/06/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:50  
PRESSURIZING 1:56  
STERILIZING 2:50  
136.4 C 222kPa 2:50  
136.3 C 224kPa 3:20  
136.4 C 223kPa 3:50  
136.4 C 224kPa 4:20  
136.3 C 221kPa 4:50  
136.4 C 225kPa 5:20  
136.4 C 224kPa 5:50  
136.4 C 224kPa 6:20  
136.4 C 224kPa 6:20  
VENTING 6:21  
AIR DRYING 7:10  
CYCLE COMPLETE 7:13

-----  
Digital Signature #  
C23C554F94543753  
-----