
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 001425
11:23 09/13/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:03
PRESSURIZING 7:06
STERILIZING 8:49
136.0 C 220kPa 8:49
136.5 C 222kPa 9:19
136.4 C 222kPa 9:49
136.2 C 219kPa 10:19
136.5 C 222kPa 10:49
136.4 C 222kPa 11:19
136.0 C 219kPa 11:49
136.2 C 220kPa 12:19
VENTING 12:19
AIR DRYING 13:22
CYCLE COMPLETE 43:22

Digital Signature #
DFF994EABACF7A8C
