

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001485  
10:29 26/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:08  
PRESSURIZING 3:10  
STERILIZING 4:18  
136.4 C 225kPa 4:18  
136.4 C 225kPa 4:48  
136.3 C 225kPa 5:18  
136.4 C 224kPa 5:48  
136.3 C 224kPa 6:18  
136.4 C 226kPa 6:48  
136.4 C 225kPa 7:18  
136.4 C 225kPa 7:48  
136.4 C 225kPa 7:48  
VENTING 7:48  
AIR DRYING 8:42  
CYCLE COMPLETE 8:48

-----  
Digital Signature #  
C7F77505CADA2BF0  
-----