
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001519
8:00 21/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:11
STERILIZING 4:11
136.5 C 220kPa 4:11
136.2 C 222kPa 4:41
136.2 C 221kPa 5:11
136.5 C 224kPa 5:41
136.4 C 223kPa 6:11
136.3 C 220kPa 6:41
136.0 C 219kPa 7:11
136.1 C 221kPa 7:41
136.1 C 221kPa 7:41
VENTING 7:42
AIR DRYING 8:32
CYCLE COMPLETE 8:34

Digital Signature #
9F32F13A6E93ECEB
