
Midmark M11 - v1.0.0
Total Cycles: 5069

Name: RITTER 1
Sterilizer ID: V2178198
11/07/2024 03:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.9 F 0.0
2:00 122.7 F 0.1
4:00 154.8 F 0.5
6:00 177.1 F 2.5
8:00 202.0 F 6.2
10:00 226.8 F 10.5
12:00 228.4 F 6.6
14:00 234.4 F 9.6
16:00 237.0 F 9.6
18:00 242.9 F 11.9
20:00 240.2 F 10.2
22:00 252.9 F 16.5
24:00 263.9 F 22.9
26:00 271.6 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.5 F 28.7
2:00 272.2 F 28.9
3:00 272.2 F 28.5
4:00 272.6 F 29.1

Min 271.5 F 28.3
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.4 F 0.0
5:00 242.7 F 0.0
10:00 215.1 F 0.0
15:00 201.4 F 0.0
20:00 243.9 F 0.0
25:00 227.1 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 26:51
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:05:29
CYCLE COMPLETE

11/07/2024 04:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____