
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006016
14:07 19/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:32
STERILIZING 3:27
136.4 C 222kPa 3:27
136.4 C 225kPa 3:57
136.4 C 223kPa 4:27
136.3 C 223kPa 4:57
136.4 C 223kPa 5:27
136.3 C 223kPa 5:57
136.4 C 224kPa 6:27
136.1 C 222kPa 6:57
136.1 C 222kPa 6:57
VENTING 6:57
AIR DRYING 7:47
CYCLE COMPLETE 11:34

Digital Signature #
02D2B35F53046BFB
