
Midmark M11 - v1.0.0
Total Cycles: 3330

Name: RITTER 1
Sterilizer ID: V2178198
06/02/2023 03:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.9 F 0.0
2:00 189.7 F 1.6
4:00 209.7 F 7.1
6:00 223.1 F 12.7
8:00 229.3 F 8.4
10:00 238.9 F 11.4
12:00 242.6 F 12.2
14:00 241.5 F 11.1
16:00 254.9 F 17.9
18:00 266.5 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.7 F 29.0
2:00 271.3 F 28.3
3:00 272.7 F 29.1
4:00 271.8 F 28.4

Min 271.3 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.1 F 0.0
5:00 206.0 F 0.0
10:00 230.8 F 0.0
15:00 223.2 F 0.0
20:00 219.3 F 0.0
25:00 218.9 F 0.0
30:00 215.1 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:58:43
CYCLE COMPLETE

06/02/2023 04:45 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____