
Midmark M11 - v1.0.5
Total Cycles: 3562

Name: RITTER2
Sterilizer ID: V2313228
04/26/2024 08:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.2 F 0.0
2:00 85.4 F 0.1
4:00 87.1 F 1.1
6:00 128.5 F 3.5
8:00 157.5 F 6.7
10:00 179.3 F 10.0
12:00 218.9 F 8.0
14:00 221.5 F 6.5
16:00 229.0 F 10.3
18:00 231.7 F 8.3
20:00 237.0 F 10.6
22:00 234.8 F 8.1
24:00 247.1 F 14.0
26:00 258.3 F 20.1
28:00 268.4 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.0 F 29.1
2:00 272.6 F 29.5
3:00 271.8 F 29.5
4:00 271.6 F 28.9

Min 271.2 F 28.9
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.1 F 0.0
5:00 222.8 F 0.0
10:00 216.7 F 0.0
15:00 228.4 F 0.0
20:00 219.9 F 0.0
25:00 229.1 F 0.0
30:00 217.0 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 29:56
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:07:44
CYCLE COMPLETE

04/26/2024 09:36 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____