
Midmark M11 - v1.0.0
Total Cycles: 891

Name: RITTER 1
Sterilizer ID: V2178198
06/14/2021 09:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.6 F 0.0
2:00 144.5 F 2.3
4:00 171.6 F 5.9
6:00 197.3 F 10.8
8:00 215.3 F 5.1
10:00 228.8 F 10.9
12:00 230.7 F 7.9
14:00 242.4 F 11.9
16:00 243.7 F 12.5
18:00 256.0 F 19.0
20:00 267.0 F 25.6
22:00 272.3 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 272.5 F 29.2
2:00 272.8 F 29.6
3:00 272.1 F 29.6
4:00 272.2 F 29.1

Min 271.6 F 29.0
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.4 F 0.0
5:00 235.5 F 0.0
10:00 229.1 F 0.0
15:00 225.7 F 0.0
20:00 230.7 F 0.0
25:00 227.1 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 22:06
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:00:28
CYCLE COMPLETE

06/14/2021 10:41 AM
ACCEPTED: X
REJECTED:
Operator: U01