
Midmark M11 - v1.0.0
Total Cycles: 4448

Name: RITTER 1
Sterilizer ID: V2178198
04/11/2024 12:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.4 F 0.0
2:00 137.2 F 1.9
4:00 152.2 F 4.9
6:00 178.6 F 8.5
8:00 206.1 F 12.8
10:00 215.9 F 5.6
12:00 225.5 F 9.9
14:00 232.9 F 9.0
16:00 238.7 F 11.4
18:00 237.3 F 9.5
20:00 250.0 F 15.6
22:00 261.2 F 21.9
24:00 271.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.0
1:00 272.6 F 29.7
2:00 272.1 F 29.1
3:00 271.8 F 29.4
4:00 272.3 F 29.4

Min 271.4 F 28.8
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 225.5 F 0.0
10:00 227.6 F 0.0
15:00 227.6 F 0.0
20:00 228.3 F 0.0
25:00 229.4 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 25:22
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:04:13
CYCLE COMPLETE

04/11/2024 01:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____