
Midmark M11 - v1.0.0
Total Cycles: 2817

Name: RITTER 1
Sterilizer ID: V2178198
01/25/2023 09:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.8 F 0.0
2:00 199.4 F 0.0
4:00 198.7 F 3.2
6:00 213.8 F 8.4
8:00 231.4 F 11.3
10:00 233.6 F 9.4
12:00 239.3 F 9.2
14:00 245.2 F 13.1
16:00 244.1 F 12.0
18:00 257.1 F 19.0
20:00 268.7 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.7 F 28.9
2:00 271.9 F 28.1
3:00 272.2 F 28.7
4:00 272.6 F 28.8

Min 271.6 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.8 F 0.0
5:00 220.8 F 0.0
10:00 238.2 F 0.0
15:00 237.0 F 0.0
20:00 225.6 F 0.0
25:00 221.7 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 21:10
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

01/25/2023 10:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____