
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002224
14:48 05/11/2024
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 5:19
STERILIZING 6:25
136.4 C 227kPa 6:25
136.0 C 222kPa 6:55
136.2 C 223kPa 7:25
136.4 C 227kPa 7:55
136.4 C 226kPa 8:25
136.0 C 223kPa 8:55
136.4 C 226kPa 9:25
136.3 C 224kPa 9:55
136.5 C 226kPa 10:25
136.4 C 227kPa 10:55
136.2 C 225kPa 11:25
136.4 C 227kPa 11:55
136.4 C 227kPa 12:25
136.4 C 225kPa 12:55
136.2 C 224kPa 13:25
136.3 C 225kPa 13:55
136.2 C 223kPa 14:25
136.4 C 226kPa 14:55
136.4 C 226kPa 15:25
136.4 C 226kPa 15:55
136.4 C 226kPa 16:25
136.4 C 226kPa 16:25
VENTING 16:25
AIR DRYING 17:19
CYCLE COMPLETE 47:20

Digital Signature #
39589EE531C687D8
