
Midmark M11 - v1.0.5
Total Cycles: 271

Name: RITTER2
Sterilizer ID: V2313228
08/09/2021 11:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.5 F 0.0
2:00 194.7 F 3.7
4:00 199.0 F 10.1
6:00 217.1 F 5.3
8:00 233.3 F 12.0
10:00 238.5 F 10.4
12:00 238.3 F 9.7
14:00 253.2 F 17.0
16:00 265.4 F 24.3
18:00 272.5 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.5 F 29.8
2:00 271.7 F 29.1
3:00 273.1 F 29.9
4:00 272.2 F 29.1

Min 271.6 F 29.0
Max 273.4 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.7 F 0.0
5:00 235.0 F 0.0
10:00 223.6 F 0.0
15:00 225.7 F 0.0
20:00 225.4 F 0.0
25:00 228.2 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 2:25
HEATING: 18:26
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 00:57:48
CYCLE COMPLETE

08/09/2021 12:17 PM
ACCEPTED: X
REJECTED:
Operator: U01