
Midmark M11 - v1.0.5
Total Cycles: 4225

Name: RITTER2
Sterilizer ID: V2313228
12/18/2024 11:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.3 F 0.0
2:00 189.0 F 2.0
4:00 190.3 F 6.6
6:00 210.6 F 12.1
8:00 227.3 F 7.8
10:00 240.5 F 13.3
12:00 239.2 F 11.0
14:00 237.8 F 9.8
16:00 251.8 F 16.5
18:00 263.4 F 23.3
20:00 272.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.6 F 29.3
2:00 271.4 F 28.9
3:00 272.5 F 29.3
4:00 271.3 F 28.8

Min 271.2 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 232.5 F 0.0
10:00 237.5 F 0.0
15:00 224.3 F 0.0
20:00 218.1 F 0.0
25:00 216.0 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 20:09
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 00:59:42
CYCLE COMPLETE

12/18/2024 12:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____