

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2277

Name: RITTER2  
Sterilizer ID: V2313228  
04/20/2023 01:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.1 F 0.0  
2:00 182.9 F 3.9  
4:00 199.8 F 8.8  
6:00 221.0 F 8.2  
8:00 231.4 F 10.4  
10:00 233.3 F 8.0  
12:00 240.8 F 10.6  
14:00 247.9 F 14.5  
16:00 260.7 F 21.8  
18:00 272.3 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.9  
1:00 272.1 F 29.6  
2:00 272.1 F 29.2  
3:00 271.2 F 28.9  
4:00 272.4 F 29.7  
  
Min 271.2 F 28.7  
Max 272.5 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.3 F 0.0  
5:00 232.1 F 0.0  
10:00 231.5 F 0.0  
15:00 227.4 F 0.0  
20:00 222.0 F 0.0  
25:00 224.9 F 0.0  
30:00 224.3 F 0.0  
Drying Complete

FILLING: 2:03  
HEATING: 18:21  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:57:38  
CYCLE COMPLETE

04/20/2023 01:58 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_